

Residual risks of polymer-based CAD/CAM materials (such as VITA CAD-Temp and CAD-Temp IS)

Possible complications and residual risks of polymer-based CAD/CAM materials in general, and not specific to VITA materials, are:

- Peeling or loss of the restoration resulting in possible choking
- Post-operative hypersensitivity
- Plaque deposits
- Irritation of the mucosa
- Gingivitis
- Intolerance and/or potential allergic reaction
- Caries in marginal gaps
- Breakage

These risks must be communicated to the patient by trained personnel (e.g., dental professionals).