

**Residual risks of polymer-based CAD/CAM materials (such as VITA CAD-Waxx)**

*Possible complications and residual risks of polymer-based CAD/CAM materials in general, and not specific to VITA materials, are:*

- Irritation of the mucosa
- Intolerance and/or potential allergic reaction
- Breakage of the material in the mouth, with possible injury to the surrounding tissue

*These risks must be communicated to the patient by trained personnel (e.g., dental professionals).*

002/11.2021