

Residual risks of zirconia CAD/CAM materials (such as VITA YZ SOLUTIONS)

Possible complications and residual risks of zirconia CAD/CAM materials in general, and not specific to VITA materials, are:

- Peeling or loosening of the restoration, resulting in possible choking
- Plaque deposits
- Post-operative hypersensitivity
- Inflammation of the mucosa
- Caries in marginal gaps
- Loosening of the abutment teeth
- Intolerance and/or potential allergic reaction
- Breakage

These risks must be communicated to the patient by trained personnel (e.g., dental professionals).

002/11.2021