

Important!

The different VITA YZ Materials have different sintering parameters!




For further information please read the instructions for use on
www.vita-zahnfabrik.com/cadcam

For VITA YZ MULTI TRANSLUCENT we recommend a temperature adjustment of +70°C for optimal results with other furnaces.

VITA

Nesting recommendation

VITA YZ® MULTI TRANSLUCENT



VITA YZ® XT Multicolor / VITA YZ® ST Multicolor

