VITA Zahnfabrik
H. Rauter GmbH & Co. KG
Spitalgasse 3
79713 Bad Säckingen
Germany
008/11.2026

The following SSCP (Summary of Safety and Clinical Performance) is applicable to the following product(s):

VITA VM®CC

Table of content

Identification of the device 2
Indications, Intended Purpose and Target populations 2
Device description 3
a) Description of the medical device(s) 3
b) Previous generations of the medical device(s) 3
c) Accessories / other products which are intended to be used with the medical device(s) 3
Possible therapeutic or diagnostic alternatives 4
Reference to harmonized standards and CS applied 5
Summary of clinical data 7
a) Clinical studies of the medical device(s) 7
b) Clinical evaluation 7
c) Post market clinical follow-up 8
d) Conclusion of clinical performance and safety of the medical device(s) 8
Suggested profile and training of users 9
Information on residual risks, undesirable effects and warnings and precautions 9
a) Residual Risks 9
b) Undesirable effects 9
c) Warnings and precautions 9

Revision History

Version	Changes
006	**New warning**
007	**New form**
008	**Formal Update on first page and updating standards**

VITA Zahnfabrik
H. Rauter GmbH & Co. KG
Spitalgasse 3
79713 Bad Säckingen
Germany
008/11.2026

Identification of the device



Device trade name	VITA VM®CC
Manufacturer	VITA Zahnfabrik H. Rauter GmbH & Co. KG Spitalgasse 3 D-79713 Bad Säckingen
Manufacturers SRN	DE-MF-000005906
BASIC-UDI-DI	++J017BC2PT
Medical device nomenclature (EMDN)	Q010699 - MATERIALS FOR THE PREPARATION OF CUSTOM-MADE DENTAL DEVICES – OTHER
Class of device	IIa
Year of first CE certificate	2007
Notified Body including identification no.	DEKRA Certification GmbH, Identification no.: 0124

Indications, Intended Purpose and Target populations

Intended purpose	VITAVM® CC products are veneer resin materials for dental prosthesis.
Indication	• Fabrication of temporary crowns and bridges • Cementation of VITA denture teeth, e.g., on model casts • Repairs • Coverage of jointed prostheses in the case of combined work
Contraindication	• Permanent veneers • In case of parafunction • Not recommended for bridge restorations without a substructure
Intended user	Dental technician, Dentist, Professional User, Rx only

Device description

a) Description of the medical device(s)

VITA VM CC is a filler-free cold curing polymer resin for extra-oral fabrication of indirect restorations and repairs. VITA VM CC is a powder in a can.

b) Previous generations of the medical device(s)

VITA VM CC is the follow-up product of VITA ZETA CC POLYMER.

c) Accessories / other products which are intended to be used with the medical device(s)

The following products can, but do not have to be used with the product. For a detailed description of how to use the products in combination, please refer to the product's instructions for use.

Name of accessory (manufacturer)	Short description
VITA VM LC SEPARATOR	Blow the model dry and block any undercuts using wax. Then immerse the model in water for five minutes and coat sufficiently with low-viscosity plaster-acrylic insulation (alginate insulation). Alternatively, the VITA VM LC SEPARATOR can also be used. Please adhere to the working instructions.
VITA VM CC LIQUID	A large measuring spoon of VITA VM CC powder (approx. 0.23 g) is mixed with four drops (approx. 0.13 g) of VITA VM CC LIQUID.

Possible therapeutic or diagnostic alternatives

Diagnostic/therapeutic alternative with conditions of use	Possible benefit/advantage and possible risks/disadvantages as far as known
VITA CAD-Temp is a therapeutic CAD/CAM alternative to VITA VM CC. In general, therapeutic alternatives are comparable composites or also polymer discs or blocks for subtractive manufacturing. Dual-curing composite materials from the cartridge for single-tooth restorations and smaller bridges are also the most common form for temporary restorations and a conceivable alternative for VITA VM CC.	Risks of these therapeutic alternatives could be allergic reactions to the adhesive cementation (or the residual monomer content present in it) when using composite where this type of cementation is necessary

Reference to harmonized standards and CS applied

Common specifications are not used for the product VITA VM LC and VITA VM LC flow. The following standards are applied for these products at VITA:

- *MDCG 2021-3 Custom-Made Devices Guideline*
- *ASTM D4332-22 Standard Practice for Conditioning Containers, Packages, or Packaging Components for Testing*
- *ASTM D5276-19 (2023) Standard Test Method for drop Test of Loaded Containers by Free Fall*
- *ASTM F1980-21 Standard Guide for Accelerated Aging of Sterile Barrier Systems and Medical Devices*
- *DIN EN 1641:2010-02 Dentistry – Medical devices for dentistry – Materials;*
- *DIN EN 22248:1993-02 Packaging Complete filled transport packages, Vertical impact test by droppinig*
- *DIN EN 62366:2021-08 Medical devices - Application 01 usability engineering to medical devices*
- *DIN EN ISO 10477:2021-02 Dentistry– Polymer-based crown and veneering materials*
- *DIN EN ISO 10993-1:2026-03 Biologische Beurteilung von Medizinprodukten - Teil 1: Anforderungen und allgemeine Grundsätze für die Beurteilung der biologischen Sicherheit im Rahmen eines Risikomanagementsystems*
- *DIN EN ISO 10993-10:2023-04 Biological evaluation of medical devices – Part 10: Tests for skin sensitization*
- *DIN EN ISO 10993-11:2018-09 Biological evaluation of medical devices– Part 11: Tests for systemic toxicity*
- *DIN EN ISO 10993-12:2021-08 Biological evaluation of medical devices – Part 12: Sample preparation and reference materials*
- *DIN EN ISO 10993-13:2024-04 Biological evaluation of medical devices – Part 13: Identification and quantification of degradation products from polymeric medical devices*
- *DIN EN ISO 10993-14:2024-04 Biological evaluation of medical devices - Part 14: Identification and quantification of degradation products from ceramics*
- *DIN EN ISO 10993-16:2024-04Biological evaluation of medical devices – Part 16: Toxicokinetic study design for degradation products and leachables*
- *DIN EN ISO 10993-17:2026-05 Biological evaluation of medical devices - Part 17: Toxicological risk assessment of medical device constituents*
- *DIN EN ISO 10993-18:2023-11 Biological evaluation of medical devices - Part 18: Chemical characterization of materials*
- *DIN EN ISO 10993-2:2023-02 Biological evaluation of medical devices – Part 2: Animal welfare requirements*
- *DIN EN ISO 10993-23:2021-10 Biological evaluation of medical devices - Part 23: Test for irritation*
- *DIN EN ISO 10993-3:2026-11 Biological evaluation of medical devices – Part 3: Tests for genotoxicity, carcinogenicity and reproductive toxicity*
- *" DIN EN ISO 10993-4:2017-12 Biological evaluation of medical devices– Part 4: Selection of tests for interactions with blood"*
- *DIN EN ISO 10993-5:2024-04 Biological evaluation of medical devices – Part 5: Tests for in vitro cytotoxicity*
- *DIN EN ISO 10993-6:2026-08 Biological evaluation of medical devices - Part 6: Tests for local effects after implantation*
- *DIN EN ISO 10993-9:2022-03 Biological evaluation of medical devices – Part 9: Framework for identification and quantification of potential degradation products*
- *DIN EN ISO 13485:2021-12 Medical devices - Quality management systems — Requirements for regulatory purposes*
- *DIN EN ISO 14971:2022-04 Medical devices - Application of risk management to medical devices*
- *DIN EN ISO 15223-1:2022-02 Medical devices – Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements*
- *DIN EN ISO 18675:2023-04 Dentistry - Machinable ceramic blanks*
- *DIN EN ISO 20417:2022-03 Information to be supplied by the manufacturer of medical devices;*

- *DIN EN ISO 4180:2020-03 Packaging - Complete filled tranport packages- Gerneral rules for compilation of performance test schedules*
- *DIN EN ISO 6872:2024-12 Dentistry - Ceramic materials*
- *DIN EN ISO 9693:2020-02 Dentistry – Compatibility testing for metal-ceramic and ceramic-ceramic systems*
- *ISO 10993-1:2025 Biological evaluation of medical devices — Part 1: Requirements and general principles for the evaluation of biological safety within a risk management process*
- *ISO 13485:2016-03 Medical devices — Quality management systems — Requirements for regulatory purposes*
- *ISO 15223-1:2021-07 Amd 1:2025-03 Medical devices - Symbols to be used with information to be supplied by the manufacturer - Part 1: General requirements*
- *ISO 15223-1:2021-07 Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied -by the manufacturer - Part 1: General requirements*
- *ISO 18675:2022-05 Dentistry - Machinable ceramic blanks*
- *ISO 20417:2021-12 Information to be supplied by the manufacturer of medical devices*
- *ISO 4180:2019-11 Packaging - Complete filled tranport packages- Gerneral rules for compilation of performance test schedules*
- *ISO 7405:2025-06 Dentistry — Evaluation of biocompatibility of medical devices used in dentistry*
- *ISO 9693:2019-10 Dentistry — Compatibility testing for metal-ceramic and ceramic-ceramic systems*
- *ISO TR 24971:2020-06 Medical devices — Guidance on the application of ISO 14971*
- *Use of International Standard ISO 10993-1, "Biological Evaluation of medical devices - Part 1: Evaluation and testing within a risk management process". Guidance for Industry and Food and Druf Administration Staff. September 8, 2023*
- *MDCG 2018-1 Rev.4 Guidance on BASIC-UDI and changes to UDI-DI*
- *MDCG 2019-16 Guidance on Cybersecurity for medical devices*
- *MDCG 2019-4 Timelines for registration of device data elements in EUDAMED*
- *MDCG 2019-5 Registration of Legacy Devices in EUDAMED*
- *MDCG 2019-7 Guidance on Article 15 MDR-IVDR on a Person responsible for Regulatory Compliance (PRRC)*
- *MDCG 2019-9 Summary of safety and clinical performance*
- *MDCG 2020:6 Regulation (EU) 2017/745: Clinical evidence needed for medical devices previously CE marked under Directives 93/42/EEC or 90/385/EEC*
- *MDCG 2020-3 Guidance on significant changes regarding the transitional provision under Article 120 of the MDR with regard to devices covered by certificates according to MDD or AIMDD*
- *MDCG 2020-7 Post-market clinical follow-up (PMCF) Plan Template - A guide for manufacturers and notified bodies*
- *MDCG 2020-8 Post-market clinical follow-up (PMCF) Evaluation Report Template - A guide for manufacturers and notified bodies*
- *MDCG 2021-1 Rev.1 Guidance on harmonised administrative practices and alternative technical solutions until EUDAMED is fully functional*
- *MDCG 2021-19 Guidance note integration of the UDI within an organisation`s quality management system*
- *MDCG 2021-25 Rev. 1/ Regulation (EU) 2017/745 - application of MDR requirements to 'legacy devices' and to devices placed on the market prior to 26 May 2021 in accordance with Directives 90/385/EEC or 93/42/EEC*
- *MDCG 2022-4 Rev. 2 Guidance on appropriate surveillance regarding the transitional provisions under Article 120 of the MDR with regard to devices covered by certificates according to the MDD or the AIMDD*
- *MDCG 2023-3 Rev. 2 Questions and Answers on vigilance terms and concepts as outlined in the Regulation (EU) 2017/745 and Regulation (EU) 2017/746*
- *MDCG 2023-7 Guidance on exemptions from the requirement to perform clinical investigations pursuant to Article 61(4)-(6) MDR and on 'sufficient levels of access' to data needed to justify claims of equivalence*
- *MEDDEV 2_7_1_rev4_en 2016-06 Guideline on Medical Devices*
- *Recommendation-NB-MED/2_5-2/Rec2_Reporting of design changes and changes of the quality system; 2.5.2 Conformity assessment procedures; Quality assurance.*

Summary of clinical data

a) Clinical studies of the medical device(s)

According to Article 61 European Medical Device Regulation, (EU) 2017/745 (MDR) the requirement to perform clinical investigations pursuant to paragraph 4 shall not apply to implantable devices and class III devices:

- which have been lawfully placed on the market or put into service in accordance with Directive 90/385/EEC or Directive 93/42/EEC and for which the clinical evaluation is based on sufficient clinical data, and
- is in compliance with the relevant product-specific CS for the clinical evaluation of that kind of device, where such a CS is available; or
- that are sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips or connectors for which the clinical evaluation is based on sufficient clinical data and is in compliance with the relevant product-specific CS, where such a CS is available.

The products under evaluation have been lawfully placed on the market and put into service in accordance with Directive 93/42/EEC and the clinical evaluation of these products is based on sufficient clinical data, and is in compliance with the relevant product-specific CS for the clinical evaluation of that kind of device. Therefore, clinical investigations are not necessary to proof clinical safety and clinical evaluation is performed.

b) Clinical evaluation

Extract from the final summary of the clinical evaluation "VITA VM CC VITA VM LC VITA Zahnfabrik H. Rauter GmbH & Co. KG" at novineon CRO GmbH June 2024/Rev.04: "[…] Based on the documentation provided by VITA, we conclude that the potential risks of the VITA VM CC, VITA VM LC and VITA AKZENT LC are acceptable residual risks for the patient and the user. The main risks, chipping or fracture of the temporal dental restorations manufactured and suspicion of toxicity effect of raw materials, are described in detail in the scientific literature. The trained dental technician and/or dental physician knows how to handle, process, and combine these raw materials to manufacture a safe and esthetic dental restoration with a good esthetic appearance. Thus, we conclude that the risks of the use of veneering materials, indications, contraindications and warnings, described in the IFU are well-documented in the published literature and the state-of-the-art, thus being known to dentists or adequately trained staff (chapter 7.4.3.2.1). When complying with all warnings and precautions, VITA VM CC, VITA VM LC and VITA AKZENT LC have an acceptable benefit-risk profile."

c) Post market clinical follow-up

PMCF studies may not be required when the medium/long-term safety and clinical performance are already known from previous use of the device or where other appropriate post-market surveillance activities would provide sufficient data to address the risks.
The indication and treatments of VITA dental medical devices are well-known clinical procedures. The basic principles of using dental material are the same since the beginning of the 20th century. The compliance to the DIN EN ISO 10477 applicable for polymer-based crown and veneering material ensures the safe use of the products also in the clinical aspect.

The VITA post market monitoring collects clinically relevant data to an extent that fulfills the requirements of the European Medical Device Regulation, (EU) 2017/745 (MDR) in order to adequately assess and confirm the safety of medical devices. Therefore, no PMCF studies are required for VITA VM®CC and therefore, also not part of this SSCP.

d) Conclusion of clinical performance and safety of the medical device(s)

VITA always ensures the clinical safety of its products, even after they have been placed on the market, by constantly updating the clinical evaluation of its medical products and monitoring them on the market in accordance with the requirements of the European Medical Device Regulation, (EU) 2017/745 (MDR) and according to MEDDEV 2.7/1 revision 4.

For VITA VM®CC it can be said that the clinical evaluation of the products clearly states the clinical safety and performance. There is no doubt about the safety and reliability of the products. They can be used safely in the manner communicated by VITA with regard to indications, contraindications, compliance with safety instructions and residual risks.

Suggested profile and training of users

VITA dental products are designed for use by professional users. This specification is made clear by the labeling of VITA products with the symbol "Rx only". The specialist users are dentists and dental technicians who have excellent prior knowledge in the use of our products due to their many years professional training and/or university education. Follow-up training is the responsibility of the expert users and is offered by VITA specifically for VITA products. This guarantees safe handling of VITA products at every point in the application process.

Information on residual risks, undesirable effects and warnings and precautions

a) Residual Risks

Information on product-specific residual risks can be found on the VITA homepage at Downloadcenter. Produktsicherheit (vita-zahnfabrik.com)

Please be aware, that these are possible complications and residual risks of the dental product group in general, and not specific to VITA materials.

These risks must be communicated to the patient by trained personnel (e.g., dental professionals).

b) Undesirable effects

There are no known undesirable effects for the products.

c) Warnings and precautions

- **While work is in progress, wear suitable protective gloves / protective clothing / safety goggles.**
- **Only perform work under an extraction unit.**
- **Dispose in accordance with official regulations as hazardous waste.**
- **Keep away from ignition sources.**



The following VITA VM CC products or accessories require hazard identification:		
VITA VM CC LIQUID (Contains: methyl methacrylate, ethylene dimethacrylate, N, N-dimethyl-p-toluidine.)	H317: May cause an allergic skin reaction. H350: May cause cancer.	
VITA VM CC POLYMER-powder (Contains dibenzoyl peroxide)	May cause allergic skin reactions.	

All these warnings and precautions can also be found in the corresponding instructions for use.